

Strengths and weaknesses per station

NAWIs (retail, medical, weighbridges):

Areas of strength (all or most candidates provided satisfactory answers):

- Markings on a NAWI.
- Tests to conduct on initial verification.
- UKCA mark.
- Accuracy classes and their intended use.
- 2016 regulations.
- MPEs.
- Medical NAWIs.

Lack of knowledge/understanding (some or most candidates provided incorrect or incomplete answers):

- Scenario question . replacement of POS on checkout scales.
- Conformity assessment procedures (e.g. B+D).
- Substitution material procedure for weighbridges.
- Compatibility of modules between indicator and load cells.
- Information found in a Type Examination Certificate.
- Who can carry out a requalification.
- 2000 regulations.
- Scale

AWIs, packaged goods:

Areas of strength (all or most candidates provided satisfactory answers):

- Packaged goods regulations, test methods

Capacity Serving Measures e.g. line / brim (some candidates required prompt).

Tests to conduct on initial verification.

"Aware" of UKCA mark.

MPEs.

Bottle top pourer not to be used for prescribed spirits.

Some candidates aware of issue with Pernod.

Some struggled with conformity assessment procedures.
Some struggled with the identification of the bulk fuel tanker dipstiares